



Inspection Certificate

Product name	Thread Plug Gage	Standard temperature	20 °C
Code No.	4130-4	Based on	ISO 1502 (1996)
Size	M4x0.7 6H	Traceable to	10339/D-K-15015-01-02/2012-09
Serial No.	0511220333		31-0830/DKD-K-06901/2012-09

Unit: mm

		Pitch Diameter	Major Diameter
GO	Specified	3.551±0.005	4.006±0.009
	Observed	3.552	4.009
NO GO	Specified	3.668±0.005	3.808±0.009
	Observed	3.665	3.809

Judgement: Passed

Inspector: *S. Yang*



WARRANTY CERTIFICATE

Inspector

22110343

INSIZE products conform to respective standards. In the event that the product fails within one year from the date of purchase, please contact us. We will repair or replace at our option, free of charge. This warranty does not apply if the products are subject to fair wear, or improper use / handling / storage / maintenance / service / repair.

INSIZE
www.insize.com

INSIZE DO (BRAZIL) LTDA.
www.insize.com.br

INSIZE INC. (USA)
www.insizeus.com

INSIZE (CHINA) CO., LTD.
www.insize.cn

INSIZE (EUROPE)
www.insize-eu.com

INSIZE (INDIA)
www.insize.in