



# Inspection Certificate

Product name	Fine Thread Plug Gage	Standard temperature	20 °C
Code No.	4139-42V	Based on	ISO 1502 (1996)
Size	M42x3 6H	Traceable to	10339/D-K-15015-01-02/2012-09
Serial No.	1504210143		31-0830/DKD-K-06901/2012-09

Unit: mm

		Pitch Diameter	Major Diameter
GO	Specified	40.067±0.007	42.016±0.014
	Observed	40.069	42.013
NO GO	Specified	40.323±0.007	Major Diameter 40.923±0.014
	Observed	40.324	40.926

Judgement: Passed

Inspector: S. Yang



## WARRANTY CERTIFICATE

### 21040657

Inspector \_\_\_\_\_

INSIZE products conform to respective standards. In the event that the product fails within one year from the date of purchase, please contact us. We will repair or replace at our option, free of charge. This warranty does not apply if the products are subject to fair wear, or improper use / handling / storage / maintenance / service / repair.

INSIZE  
www.insize.com

INSIZE DO (BRAZIL) LTDA.  
www.insize.com.br

INSIZE INC. (USA)  
www.insizeus.com

INSIZE (CHINA) CO., LTD.  
www.insize.cn

INSIZE (EUROPE)  
www.insize-eu.com

INSIZE (INDIA)  
www.insize.in