

Test Report

Test bench serial number: 0212
Transducer type: WIKA S-20
Ordering number: -
Range: 0-1000 bar
Output signal: 4-20 mA
Power supply: -
Transducer serial number: 1A02FLUNQJL
Digital indicator: PLC+PC
Line: B
Station: 2

Test report n. **04_10_2024_TR_00000419**
Customer: JC FABRICA DE VALVULAS S.A.U

Certificate

BIT02B - 924687 - LAT 093 170719P

Climbing measure

Pressure	Measure	Error limits measured (%)
0.00	0.00	0.00
175.00	175.00	0.00
350.00	350.00	0.00
525.00	525.00	0.00
700.00	700.00	0.00

Downhill measure

Pressure	Measure	Error limits measured (%)
700.00	700.00	0.00
525.00	525.00	0.00
350.00	350.00	0.00
175.00	175.00	0.00
0.00	0.00	0.00



Date: 04/10/2024

Firma HPSC:

